

## PRODUCT RECALL NOTICE



Himalayan Regulator (Standard)



Himalayan Regulator (Elite)

### 1 RISK.

- 1.1 There is a risk that under certain climatic conditions (high humidity levels) the regulator can be affected which can cause the Pressure Relief Valve (PRV) to activate, resulting in the rapid loss of the cylinder contents.

### 2 PRODUCTS

- 2.1 The affected products are listed below:

Item	Description	Part No	Revision No
1	Himalayan Regulator (Standard) 4-lpm	102217	02
2	Himalayan Regulator (Standard) 6-lpm	103021	01
3	Himalayan Regulator (Standard) 8-lpm	107682	01
4	Himalayan Regulator (Elite) 4-lpm	102900	02
5	Himalayan Regulator (Elite) 6-lpm	103009	02
6	Himalayan Regulator (Elite) 8-lpm	107680	01

### 3 REMEDY

- 3.1 Customers should stop using the items listed above and return them to Summit Oxygen for modification immediately.

## **4 CONTACT DETAILS**

4.1 For further information please contact Summit Oxygen by any of the methods below:

4.1.1 E-mail: [neil.greenwood@summitoxygen.com](mailto:neil.greenwood@summitoxygen.com)

4.1.2 Telephone (UK): +44 (0) 1252 625802

4.1.3 Mobile (UK): +44 (0) 788 400 6610

4.1.4 Mobile (Nepal): +977 98010 53707